

A-Tech International Corporation
22817 Ventura Blvd. suite 588
Woodland Hills, CA 91364
Telephone: (818) 225-0888
Fax: (818) 225-0085

**Certificate of Compliance for Radioactive Materials Package
D.O.T. Specification 7A, Type A**

1. Package Identification: Common name: Single of Multiple Dose Shipping System Model T717-S (with 11 Biodex Pigs), Syringe w/needle or without needle - Ground or Air Shipment.

2. PREAMBLE

2a. This certificate is issued to satisfy section 178.350 including sections 171.2, 173.403, 173.410, 173.411, 173.412, 173.415, 173.415 (a), 173.431, 173.465, 173.466 and 178.608 of the Department of Transportation Hazardous Materials Regulations (49 CFR).

2b. The packaging and contents described in item 4 below meet the safety standards and were tested according to strict recommendations for general performance testing from:

IATA Dangerous Good Regulations, 56th Edition, Section 5.0.4.3, 10.5, and 10.6.1 thru 10.6.3.5. 49 CFR as listed in item 2a above.

2c. This certificate does not release the consignor from complying with any requirements from the U.S Department of Transportation or any other applied regulatory agencies, including the governments of any countries through which the package will be transported.

3. Descriptions of Packaging and Authorized contents, other Conditions, And References:

PIGs :11 pieces of Biodex PIG Model 001-280,001-281,001-282
Manufacturer: Biodex
Material: Lead and Polymer
Dimensions: 1.94" (Dia) x8.52 " (H)

(Biodex is trademark of Biodex Medical Systems)

Syringe: 1- Becton , Dickinson : 3,5,8,10cc. plastic sytines

2- Duopross 3,5,10cc Retractable Safety Syringes

Needles: 1- Becton, Dickinson Company

2- Duopross

Shipping Bag: Model T-717-S Material: Nylon, PVC and Polymer laminate,
Size; 11 in (W) x 6 in(Deep) x 9.8 in (H).

Foam: Material: Polymer, A)Top foam T712-1 : size 10.4 in X 9.25in X
1.5 in. B) Bottom foam T714-1: size 10.8 inX 9.5in X 2.15 in.

Authorized Contents:

Normal form Radioactive Material in Type A quantities.

Specification and Restrictions:

Ground or Air transportation.


4. This certificate is issued on basis of the test results, safety analysis reports of the package design.

5. Test Data File Name: Single and Multiple Dose Shipping system


6. CERTIFICATE

This document is to certify that to the best of our knowledge all methods, packaging ,design and materials for building this package comply with the standard described in items 2a,2b,2c, herein. Any deviation from these conditions must be approved prior to use.

Approved:



Liu Li Hua
Manufacturer General Manager
and Senior Engineer .



Yao Su Sheng
Manufacturer Vice General Manager
and Senior Engineer.



Zhang Yong Ping
Manufacture Manager and Senior Engineer.

Date:

