


BIODEX

Biodex Medical Systems, Inc.
20 Ramsey Road, Shirley, New York, 11967-4704

DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL DIRECTIVE(S) AND NATIONAL REGULATION OF SWEDEN	93/42/EEC as amended by 2007/47/EC : LVFS 2003:11
DATE OF ISSUE	September 25, 2018
TYPE OF EQUIPMENT	Radionuclide Dose Calibrators
BRANDNAME	Atomlab 500
MODEL NUMBER(s)	086-330, 086-331, 086-332, 086-336
SERIAL NUMBER	SAMPLE
TYPE OF EQUIPMENT	Radionuclide Dose Calibrator Accessories & Replacement Parts
BRANDNAME	Atomlab
MODEL NUMBER(s)	008-015, 075-596, 086-241, 086-242, 086-243, 086-338, 086-423, 086-435, 086-509
STANDARD(s) TO WHICH CONFORMITY IS DECLARED	EN 60601-1 Ed 3 EN 60601-1-2 2007 UL 60601-1: 2003 CAN/CSA C22.2 No.:601-1-M90
BIODEX CERTIFIED QUALITY MANAGEMENT SYSTEM	ISO 13485:2003 Certificate #9060-5-03
CLASS	I Measuring Annex V
QUALITY SYSTEM REGISTRAR	INTERTEK SERVICES
E.C. CERTIFICATE # and Expiration Date	41312458-01 July 20, 2023
CE IDENTIFICATION #	0413 
ISSUED BY MANUFACTURER: Biodex Medical Systems, Inc. 20 Ramsey Road Shirley, New York 11967-4704 USA	
I, the undersigned, hereby declare that the equipment specified above conforms to the Directive(s) and Standard(s) as specified.	

Authorized European Representative



signature

Clyde Schlein
Vice President, Regulatory Affairs & Compliance



EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands

#CD-CC-135 rev F