

Biodex Medical Center

55 red st
Shirley, NY 11967

High Voltage Detail Report

Printed on: 05/08/2017 12:57:33 PM

Detector High Voltage Information

Date	: 05/08/2017 9:23:52 AM	Voltage	: 1,240.70
Detector S/N	: 2007002	662 keV Channel	: 52.00
Staff	: C S	Time (sec)	: 71.00

Status: Pass

Comments :

Signature : _____

Date Signed : _____